

A Piston Finger Device for Restoring the Motor Function of Chronic Plegic Fingers: Analysis of the Piston Finger Technique

Mengsu Wang*

*Department of Electrical and Mechanical Engineering, Nagoya Institute of Technology,
Gokiso-cho, Syowa-ku, Nagoya, Aichi 466-8555, Japan*

Hirofumi Tanabe†

*Department of Health Sciences, Shonan University of Medical Sciences,
16-48 Kamishinano, Totsuka-ku, Yokohama, Kanagawa 244-0806, Japan*

Kenji Ooka*

*Department of Electrical and Mechanical Engineering, Nagoya Institute of Technology,
Gokiso-cho, Syowa-ku, Nagoya, Aichi 466-8555, Japan*

Yoshifumi Morita*

*Department of Electrical and Mechanical Engineering, Nagoya Institute of Technology,
Gokiso-cho, Syowa-ku, Nagoya, Aichi 466-8555, Japan*

*E-mail: m.wang.188@nitech.jp, hirofumi.tanabe@sums.ac.jp, k.oooka.446@nitech.jp, morita@nitech.ac.jp
<http://www.nitech.ac.jp/eng/index.html>*

Abstract

A piston finger technique (PFT) was developed for restoring the motor function of chronic plegic fingers, with good treatment outcomes, including reduced spasticity and improved muscle shortening. Before developing a piston finger device (PFD) simulating the PFT, we analyzed the PFT by using a motion capture system. The motion of a chronic plegic index finger was investigated after administering two types of PFT-based treatment. The ranges of motion of the proximal interphalangeal and metacarpophalangeal joints of the index finger, and the vibration width and frequency during treatment were measured. These results will be useful for developing a PFD.


Keywords: Piston finger technique, Plegic finger, Spasticity, Motion analysis, Piston finger device

1. Introduction

The number of patients with cerebrovascular disease in Japan in 2014 has reached 1.179 million.¹ The main sequelae of cerebrovascular disease is hemiplegia, and approximately 80% of patients with chronic hemiplegia have chronic finger paralysis. Among the symptoms of hemiplegia, the decreased motor function of the fingers is the most difficult to recover. Owing to hand paralysis, nursing care is needed in many situations in daily life

and many patients are forced to give up self-sustaining living.

In hand rehabilitation therapy, the therapist performs treatment by moving the patient's fingers passively. However, because of the lack of therapists, many patients do not receive adequate treatment. In view of this situation, we aimed to develop a rehabilitation device to support therapists.



Currently, in Japan, a treatment method has been used for therapeutic stimulation and motion reconstruction of paralyzed limb according to upper motor neuron disorders such as stroke and spinal cord injury by therapeutic electrical stimulation and functional electrical stimulation.² In the development of equipment supporting finger rehabilitation, a perception support system has been developed that allows biofeedback transfer from the paralyzed side motion to the non-paralyzed side so that awareness can be directed to the paralyzed side movement sensation through the non-paralyzed side.³

In other countries, a robotic interface for training hand and finger functions has been developed. It is a cable-actuated rehabilitation system where each finger is attached to an instrumented cable loop that allows force control and a predominantly linear displacement.⁴

However, only a few studies have been conducted on the validation and motion analysis of manual treatment methods. Moreover, only few research studies have been conducted to develop rehabilitation devices simulating the treatment method.

On the other hand, the second author developed a piston finger technique (PFT) for restoring the motor function of chronic plegic fingers and has achieved




Fig. 2. Electromyograms of the flexor digitorum superficialis muscle of a plegic index finger.

good treatment outcomes, including reduced spasticity and improved muscle shortening.⁵


In this study, we performed myoelectric measurement to demonstrate the effectiveness of PFT for a paralyzed index finger. We also analyzed the finger motion with the PFT by using a motion capture system to clarify the finger joint angles and the frequency during the repeated motion by PFT.

2. PFT and Its Treatment Effect

2.1. What is PFT?

In general, when stretching treatment is applied to fingers with spasticity, the spastic muscle is temporarily improved. However, the spasticity often occurs again when the patient bends the affected finger. This is one of the troubles of many therapists engaged in the treatment of hemiplegia.

PFT is a technique for improving adhesion of soft tissues and spasticity. In the PFT, the therapist quickly repeats flexion and extension of the patient's finger joints. The PFT can decrease flexor spasticity in a short time and improve muscle contraction. The effect lasts longer than that by stretching treatment. Treatment with the PFT has two patterns as shown in Fig. 1.⁵



(a) Pattern 1. (b) Pattern 2.
Fig. 3. Marker attachment positions and vectors defined for computation of joint angles.

In Pattern 1, the therapist, with one hand, holds and fixes the patient's plegic thumb with full abduction, and with the other hand, holds and moves the distal phalange of the patient's index finger. From the start positions, the metacarpophalangeal (MP) and proximal interphalangeal (PIP) joints are bended and extended, respectively. Then, the metacarpophalangeal (MP) joint is bended and extended quickly and repeatedly, while the proximal interphalangeal (PIP) joint is extended and bended quickly and repeatedly. PFT treatment is applied to the index, middle, and ring fingers individually or together. After PFT treatment with Pattern 1, the flexor spasticity of the four fingers diminishes.

In Pattern 2, the start positions of the MP, PIP, and distal interphalangeal (DIP) joints are extended fully. Then, the MP joint is bended and extended quickly and repeatedly, while keeping the DIP and PIP joints in the extended position. After PFT treatment with Pattern 2 the MP joints of four fingers can bend to the vicinity of the end feel within the range without feeling pain.

2.2. Treatment effect of PFT

To verify the therapeutic effect of the PFT treatment, the therapist applied the PFT treatment to the paralyzed index finger of a male adult in his forties. The paralytic side was the left. Nine years had passed since the onset of stroke.

After treatment, the Modified Ashworth Scale (MAS) scores of the extension of the patient's plegic index finger increased from 1 to 4. This means that




Fig. 4. MP and PIP joint angles.

spastic paralysis was reduced. Electromyograms of the flexor digitorum superficialis muscle of the patient's plegic index finger before and after treatment are shown in Fig. 2. Figure 2 shows that the abnormal muscle tone of the paralyzed finger during MAS testing by the therapist. The abnormal muscle tone was markedly reduced after PFT treatment. This means that spasticity was reduced by PFT and that the therapeutic effect was confirmed.

3. Analysis of the PFT

We analyzed the patient's finger motion during PFT treatment by using the VICON motion capture system (Vicon Motion Systems Ltd.).

3.1. Motion measurement

In Pattern 1, as the DIP joint was kept fully extended, we measured the PIP and MP joint angles of the index finger as shown in Fig. 3(a). Six markers consisting of two markers between the DIP and PIP joints, two markers between the PIP and MP joints, and two markers between the MP joint and the wrist were attached on the center line of the middle phalanges, proximal phalanges, and metacarpals of the index finger, respectively. Vectors v_{11} , v_{12} , v_{13} , and v_{14} were defined as shown in Fig. 3(a).

In Pattern 2, as the DIP and PIP joints were kept fully extended, we measured only the MP joint angle of the index finger as shown in the Fig. 3(b). Four markers consisting of one marker between the DIP and PIP joints, one marker between the PIP and MP joints, and




Fig. 5. Time history of PIP joint angle in Pattern 1.




Fig. 6. Time history of MP joint angle in Pattern 1.




Fig. 7. Time history of MP joint angle in Pattern 2.

Table 1. Summaries of the joint angles in Patterns 1 and 2. (Start and End: start and end positions of the repeated motions, Diff.: difference between Start and End, x ($y - z$), x : mean value, y : maximum value, z : minimum value)

		Pattern 1 (degrees)	Pattern 2 (degrees)
PIP joint angle	Start	155.4 (135–178)	
	End	86.7 (73–96)	
	Diff.	68.7 (39–105)	
MP joint angle	Start	172.6 (168–177)	141.4 (136–148)
	End	140.4 (135–145)	109.5 (102–119)
	Diff.	32.2 (23–42)	31.9 (17–46)

two markers between the MP joint and the wrist were attached on the center line of the middle phalanges, proximal phalanges, and metacarpals of the index finger, respectively. Vectors v_{22} and v_{23} were defined as shown in Fig. 3(b).

3.2. Computation of joint angles

We computed the joint angles of the index finger during treatment. For this purpose, we measured the three-dimensional coordinates of the markers with VICON and computed the joint angles from the vectors shown in Fig. 3 (a) and (b). The PIP and MP joint angles were defined as shown in Fig. 4.

In Pattern 1, the PIP and MP joint angles were denoted by θ_{PIP1} and θ_{MP1} , respectively. The angles were computed by applying vectors v_{11} , v_{12} , v_{13} , and v_{14} to the following equations:

$$\theta_{PIP1} = \cos^{-1} \frac{v_{11} \cdot v_{12}}{|v_{11}| |v_{12}|}, \quad \theta_{MP1} = \cos^{-1} \frac{v_{13} \cdot v_{14}}{|v_{13}| |v_{14}|}. \quad (1)$$

In Pattern 2, the MP joint angle was denoted by θ_{MP2} . The angle was computed by applying vectors v_{22} and v_{23} to the following equation:

$$\theta_{MP2} = \cos^{-1} \frac{v_{21} \cdot v_{22}}{|v_{21}| |v_{22}|}. \quad (2)$$

4. Analysis of the Results of the PFT

We analyzed the patient's finger motion during treatment by the therapist for 10 seconds. The time histories of the PIP and MP joint angles during treatment of Patterns 1 and 2 are shown in Figs. 5 to 7. The summaries of the mean, maximum, and minimum values of the joint angles at the start and end positions of Patterns 1 and 2 are shown in Table 1.

In Pattern 1, we analyzed the joint angles in Figs. 5 and 6 from 2 to 10 sec. We found that the PIP joint angle was moved by the therapist between 86.7° and 155.4°, and the MP joint angle was moved between 140.4° and 172.6°. Moreover, the frequency of the repeated motion was found to be 4.25 Hz.

In Pattern 2, we analyzed the joint angle in Fig. 7 from 0 to 10 sec. We found that the MP joint angle was moved between 109.5° and 141.4°. Moreover, the frequency was found to be 3.6 Hz.

5. Conclusion

The therapeutic effect of PFT was confirmed by the comparison of MAS scores and electromyograms before and after treatment. The motion analysis revealed the joint angles and frequencies of finger motion during PFT treatment. Future studies are needed to develop the specifications of a PFT treatment apparatus.

References

1. Ministry of Health, Labor and Welfare, The overview of the patient survey in 2014. <http://www.mhlw.go.jp/toukei/saikin/hw/kanja/14/index.html> (in Japanese)
2. T. Matsunaga, The NESS H200 Hand Rehabilitation System for upper extremity paralysis [in Japanese], in *Bull. Jpn. Soc. Prosthet. Orthot. Educ. Res. Dev.*, **29**(2) (2013) 80–82.
3. H. Iwata, K. Enjyoji, R. Komagata, S. Sugano, Somatic sensation-biofeedback system for paralyzed-finger rehabilitation: development of a self-rehabilitation assisting device simultaneously cross-displaying somatic sensation of fingers [in Japanese], in *Proc. the 2008 JSME Conf. on Robotics and Mechatronics*, 2008, 1P1-D14(1)-(2).
4. R. Kato, S. Nakagawa and H. Yokoi, Development of EMG-controlled Powered Exoskeleton for Hand Rehabilitation, in *Precision Engineering Society Academic Lecture Papers Collection*, (2011), pp. 499–500.
5. L. Dovat, O. Lambercy, R. Gassert, T. Maeder, T. Milner, T. C. Leong and E. Burdet, HandCARE a cable-actuated rehabilitation system to train hand function after stroke, *IEEE Trans. Neural. Syst. Rehabil. Eng.* **16**(6) (2008) 582–591.
6. H. Tanabe, Effective approach to reconstruct central nervous system diseases: Tanabe therapy [in Japanese]. (Human Press, Tokyo, 2016).